


## INSTRUCTIONS FOR USE

### CRIMPABLE STOP

#### 1. Product Family & Part Numbers

Product Family	Product Part Numbers
Crimpable Stop  	906.0005 (10 Pack)

#### 2. Intended Use

Crimpable Stops are intended to be fixated to orthodontic archwires to maintain correct wire positioning. They are designed to fit rectangular or round wires up to .021 X .025 inches. The devices are for Professional Use Only by dental professionals in the recommended indications.

Devices are supplied:

- Non-sterile
- Designed for single-use
- For use by dentists and orthodontists only

WCT devices are intended to be used in conjunction with other orthodontic devices. WCT orthodontic appliances are used to aid in the movement of natural teeth in patients with malocclusion during orthodontic treatment.

#### 3. Indications for use

The Crimpable Stop is indicated to be placed on archwire, when correct wire positioning needs to be maintained during orthodontic treatment.

#### 4. Contraindications

- Patient's inability or unwillingness to cooperate/or follow the treatment plan.
- Known allergies to any of the components or materials in the system (see Table1).
- Any illness and/or underlying conditions that preclude orthodontic treatment.
- Any existing tooth root resorption.
- Any existing decalcification of the tooth enamel.
- Patients with deficient oral hygiene.




## 5. Materials

Stainless Steel MIM-17-4PH per ASTM B883. Material composition per Table 1 below.

**Table 1**

Stainless Steel MIM-17-4PH composition per ASTM B883-19		
Chemical Name	Min. (weight %)	Max. (weight %)
Iron (Fe)	Balance	Balance
Nickel (Ni)	3	5
Chromium (Cr)	15.5	17.5
Carbon (C)	-	0.07
Copper (Cu)	3	5
Silicon (Si)	-	1.0
Manganese (Mn)	-	1.0
Niobium; Tantalum (Nb+TA)	0.15	0.45
Other(s)	-	1.0

## 6. Warnings and Precautionary Measures

	Federal law restricts this device to the sale by or on the order of a licensed orthodontist.
	The Crimpable Stop is designed for single patient use, only, by a professional or on the order of an orthodontist or a dentist.  There is a risk of cross-contamination between patients if re-used.
	This product contains nickel and chromium, and should not be used for individuals with known allergic sensitivity to these metals. Prior to use, patients should be counseled on the materials contained in the device, as well as the potential for allergy/hypersensitivity to these materials.
Do not use this product on patients with known allergies to any of the materials used in this device. Immediately remove the device in the event of an allergic reaction.	
Follow all regional and national standards regarding the use of orthodontic appliances.	
Do not use any products which are damaged, or do not comply with the labeling specifications.	
MRI Safety Information – The Crimpable Stop has not been evaluated for safety and compatibility in the MR environment. It has not been tested for heating, migration, or image artifact in the MR environment. The safety of the device in the MR environment is unknown. Scanning a patient who has this device may result in patient injury. Always inform the MRI Technician of the presence of orthodontic devices.	

If, in relation to the use of a Crimpable Stop, a death or serious deterioration of health has occurred, this should be reported to the manufacturer and the competent authority of your country.

## 7. Patient Information

- There is no information available which would preclude the use of commonly available oral healthcare products.
- Chewing hard foods can cause appliances to be damaged.
- Some sports may cause damage to orthodontic appliances, and their presence may increase the risk of harm in the event of certain sports-related injuries.
- When participating in sports, always wear appropriate mouth and/or bracket guards as recommended by the dental professional treating the patient.
- Always inform the MRI or radiology staff that braces are in place before any procedure to ensure appropriate measures are taken for the procedure.

## 8. General Information for the Dentist/Orthodontist

- As part of developing a treatment plan, and before appliance placement, assess the need for interdisciplinary coordination with other professionals, such as speech pathologists, otolaryngologists, physicians, dentists and/or orthodontists.
- Orthodontic training in standard procedures will determine appropriate instruments for use during appliance placement and removal.
- Oral hygiene is of particular importance for immunocompromised patients. Closely monitor oral hygiene on immunocompromised patients.
- Assess whether further orthodontic treatment is advisable in the presence of root resorption.
- Crimpable Stops will be fastened (crimped) to the orthodontic archwire using an appropriate crimping plier.
  - Place Crimpable Stop at the desired location on archwire.
  - Use appropriate crimping plier and apply sufficient force to the Crimpable Stop until the appliance is firmly fixated.
- The device is provided nonsterile, nor is it intended to be sterilized by the user. The industry standard of care for orthodontic devices is that the devices are provided to the orthodontist in protective packaging, removed from the packaging for use on the patient. The devices are not handled nor sterilized prior to use.
- There is no information available that would preclude the use of commonly available oral healthcare products like toothbrush, toothpaste, water pick, mouth wash, etc.


## 9. Disposal (if applicable)

- Disposal of all orthodontic appliances must follow regional and national regulations.


### 10. Storage and Handling for medical devices (if applicable)

- The device should be stored in a dry environment under ambient conditions.

### 11. Name and address of labeler

	<p><b>World Class Technology</b>          1300 NE Alpha Dr.          McMinnville, OR 97128          USA</p>
---	---

### 12. Name, address and number of Notified Body





	<p><b>TUV Rheinland LGA Products GmbH</b>          Tillystrasse 2, 90431 Nurnberg, Germany          +49 221 808-1371          Notified Body No.: 0197</p>
---	---

### 13. Name, address, and number of Authorized Representative





	<p><b>MDI Europa</b>          Langenhagener STR.71 30855          Langenhagen, Germany          +49-511-3908-9530          SRN: DE-AR-000006218</p>
---	---




### 14. Explanation of Symbols

The following are per ISO 15223-1 (References as indicated).

Symbol Standard Reference	SYMBOL TITLE – Explanatory Text
 Ref. 5.1.1	<p><b>Manufacturer:</b>            Indicates the medical device manufacturer.</p>
 Ref. 5.1.2	<p><b>Authorized Representative:</b>            Indicates the Authorized representative in the European Community.</p>
 Ref. 5.1.5	<p><b>Batch Code:</b>            Indicates the manufacturer's batch code so that the batch or lot can be identified.</p>
 Ref. 5.1.6	<p><b>Catalogue or model number:</b>            Indicates the manufacturer's catalogue number so that the medical device can be identified.</p>

The following are per ISO 15223-1 (References as indicated).

Symbol Standard Reference	SYMBOL TITLE – Explanatory Text
 Ref. 5.2.7	<b>Non-sterile:</b> Indicates a medical device that has not been subjected to a sterilization process.
 <b>Single Use</b> Ref. 5.4.2	<b>Do not re-use/single patient use:</b> Indicates a medical device that is intended for one use, or for use on a single patient during a single procedure.
 Ref. 5.4.3	<b>Consult instructions for use:</b> Indicates the need for the user to consult the instructions for use.
 Ref. 5.7.7 MDR 2017/745 Annex 1 23.2(q)	<b>Medical Device:</b> Indicates that the device is a medical device.

Symbols Not Derived from Standards	
 21 CFR 801.109	<b>Prescription Only:</b> CAUTION: U.S. Federal law restricts this device for sale by or on order of a licensed dentist or physician.
 Contains Nickel and/or Chromium	<b>Product contains Nickel and/or Chromium</b> FDA 21 Part 872 Sec. 872.3710 Base metal alloy.
 MDD 93/42/EEC Annex II; MDR 2017/745 Annex V	<b>European conformity:</b> European conformity (CE) mark with Notified Body identification number.